



CareFlex

FLEXIBILITY IN A WORLD
OF MIXED ABILITY

The Modular Range User Guide



User Guide

Dear Client

Your new Modular chair has been designed to provide superior comfort and to be easy to use. The chair has many options so it can be adapted to meet your every seating need.

Please ensure that you read this concise user guide, especially the warnings and safety tips. We suggest that you keep this guide in the side pocket alongside the handset so that it is available for future reference.

Tip for Operation

Because of safety requirements you cannot operate the chair's footrest when the chair is tilted forwards. You have to return the chair to the neutral position before you can extend the footrest.

Maintenance and Cleaning

Regularly clean your chair with a vacuum cleaner and wipe with a clean damp cloth. If necessary dab stains with absorbent paper towels before cleaning with warm soapy water. **Do not use** aggressive cleaning agents and take care not to rub the fabric when cleaning. Always rinse with clear water. A detailed spot and stain removal guide is available on request.

Handset Buttons

Backrest adjustment (blue buttons)

Forward

Backward

Footrest adjustment (white buttons)

Up

Down

Tilt adjustment (red buttons)

Stand up / tilt forward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Sit down / tilt backward

Installation

- The Modular is meant for indoor use, in a dry environment
- The Modular requires a 220V / 240V power supply
- The working voltage of the Modular transformer is 24V
- The power supply to the motors is switched off automatically when the handset is not in use
- Check that the power cable is not pulled too tight and does not create a tripping hazard
- If your chair is not going to be used for an extended period of time (e.g. during holidays) it is recommended that you unplug the transformer
- Ensure that there is sufficient space in front and behind the chair to extend the legrest and recline the backrest fully
- Do not position the chair too close to an open fireplace or heater
- Should the handset not operate, first check that the transformer is properly plugged into the socket. If it still does not work contact your supplier
- For any special maintenance or adjustment not mentioned in this user guide contact your supplier

Safety tips

- Always grip the plug, not the cable, when disconnecting from the mains socket
- The Modular is intended for single person use
- Never sit on the armrest or legrest
- When operating the handset ensure that there is nothing near the moving parts of the chair
- Do not allow children to play with the handset
- Never put anything under the chair
- Should an object get stuck in the chair or the base, remove the plug from the mains socket before attempting to remove it
- Take care to use little or no liquid when cleaning the floor on which the chair is standing
- If you accidentally spill liquid on your chair, first remove the plug from the mains socket. Then check whether the electrical components are dry. If not, carefully wipe them with an absorbent cloth and wait until completely dry before plugging back in
- Never open the housing of the electrical components



Warning!

Do not attempt any repair work yourself – always contact your authorised Modular Range supplier



CareFlex

FLEXIBILITY IN A WORLD
OF MIXED ABILITY

The Modular Range User Guide

For more information please contact:

www.careflex.co.uk

The Modular meets the following EC directives:

- EMC Directive, 89/336/EEC, as amended
- Low Voltage Directive, 73/23/EEC, as amended recently

CareFlex® and the CareFlex waves logo are all registered trademarks of CliniMed (Holdings) Ltd